



## CLARIFICATIONS

According to the Timetable in Tender dossier form 04 INSTRUCTIONS TO TENDERERS, PUBLICATION REF.: 1328 / Mining and Metallurgy Institute Bor / 78361/15.09.2015 / 02, Published in TED website: <http://ted.europa.eu>: 2015/S 229-417239, Contract title: Supply of Research Equipment for the POCAL Project, we, as Contracting Authority, issue the following clarifications:

Lot No: 2

Lot title: DTA-TGA-DSC devices

Question 1:

Please clarify following facts in Technical specifications for Lot 2 (DTA-TGA-DSC devices):

In Specifications required is written Platinum crucibles and covers 380 ml.

Our questions are: is it typographical error, where ml stands instead  $\mu\text{l}$ ?

Will you accept offer specified like: Platinum crucibles and covers  $> 300 \mu\text{l}$  ?

Answer 1:

This is typographical error. Instead ml should be  $\mu\text{l}$ , i.e. platinum crucibles and covers 380  $\mu\text{l}$ . This capacity implies a minimum required value.

(Published on the Contracting Authority web site on March 9<sup>th</sup> 2016)